



Department of Safety and Professional Services
Division of Industry Services
Plumbing Product Review
4822 Madison Yards Way
P.O. Box 7162
Madison, Wisconsin 53707-7162
Phone 608-266-2112
Web <http://dsps.wi.gov>
Email dsps@wisconsin.gov
TTY: Contact Through Relay

Governor Tony Evers Dawn Crim, Secretary

August 22, 2019

CANTEL MEDICAL – CROSSTEX INTERNATIONAL
PABLO MARTINEZ
6789 W. HENRIETTA ROAD
RUSH NY 14543

Re: Description: HEALTH CARE PLUMBING APPLIANCE
Manufacturer: CANTEL MEDICAL – CROSSTEX INTERNATIONAL
Product Name: VISTACLEAR DENTAL WATERLINE TREATMENT SYSTEM
Model Number(s): 1000, 1000-23-C, 1000-34-C, 1000-46-C, 1000-47-C, 1000-48-C and 1000-28-HP
eSLA ID: TBD-RNW35

The specifications and/or plans for this plumbing product have been reviewed and determined to comply with chapters SPS 382 through 384, Wisconsin Administrative Code, and Chapters 145 and 160, Wisconsin Statutes.

The Department hereby issues an approval based on the Wisconsin Statutes and the Wisconsin Administrative Code. This approval is valid until the end of August 2024.

This approval supersedes the approval issued on August 15, 2014 under product file number 20140246.

This approval is contingent upon compliance with the following stipulation(s):

- This product is approved as acceptable form of backflow protection of the water supply serving dental units.
- The backflow protection devices in this product must be tested at least once per year. The testing must be done in accordance with the manufacturer's testing requirements as noted in their "Annual Backflow Preventer Valve Testing" procedures. Copy of the test results should be recorded and kept at the installation site. Any repairs or replacement of the backflow preventers should be noted on test report.

The department is in no way endorsing this product or any advertising and is not responsible for any situation which may result from its use.

Sincerely,


Glen W. Schlueter
Plumbing Product Reviewer
Department of Safety and Professional Services
Division of Industry Services
Bureau of Technical Services
(608) 267-1401 Phone
glen.schlueter@wi.gov E-mail