

# How to Use Team Vista™ Dental Waterline Cleaner

## VistaClean™ Irrigant Solution Concentrate by Hu-Friedy

## VistaTab™ Dental Waterline Cleaner Tablets by Hu-Friedy

Team Vista is a convenience combo pack that contains two different products. VistaClean is a general-purpose irrigant and cleaner that provides lubrication as an irrigant while also protecting lines and system components from deposits and scale. VistaTab is designed for the cleaning and control of microbial contaminants in dental unit waterlines. VistaClean is an FDA cleared irrigant and VistaTab is an EPA registered antimicrobial in all states.

**INITIAL CLEANING:** For first time users – Under the DIRECTIONS FOR USE section on the reverse side of this page, follow the “INITIAL TREATMENT” steps for all water delivery systems to make certain lines are cleaned prior to beginning “Daily” protocol.

### Team Vista Instructions For Dental Bottle System Users

DAILY	MONTHLY**																		
<ol style="list-style-type: none"> <li>Fill bottle to normal capacity with clean water.</li> <li>Add the correct number of VistaClean Irrigant Solution Concentrate drops based on the size of the bottle as shown below: <table border="1" data-bbox="203 961 760 1192"> <thead> <tr> <th>Bottle Capacity</th> <th>Min Drops</th> <th>Max Drops</th> </tr> </thead> <tbody> <tr> <td>500 ml</td> <td>1</td> <td>2</td> </tr> <tr> <td>750 ml</td> <td>1</td> <td>3</td> </tr> <tr> <td>1.0 Liter</td> <td>1</td> <td>3</td> </tr> <tr> <td>1.5 Liters</td> <td>1</td> <td>4</td> </tr> <tr> <td>2.0 Liters</td> <td>2</td> <td>5</td> </tr> </tbody> </table> </li> <li>Return the bottle to manifold on the dental unit. The water system is now ready for use with patients.</li> </ol> <p><i>Note: Always start with a fresh batch each time the bottle is removed and filled. Use the above formula. If water remains in the bottle at the end of the day, the bottle may be left attached to system manifold and used the following day. The bottle containing water and VistaClean drops may be left attached to the dental unit manifold when not in use to keep system sealed. If the water foams or patients notice an astringent taste, reduce the number of drops used when making each batch.</i></p>	Bottle Capacity	Min Drops	Max Drops	500 ml	1	2	750 ml	1	3	1.0 Liter	1	3	1.5 Liters	1	4	2.0 Liters	2	5	<p><i>Preparation:</i></p> <ol style="list-style-type: none"> <li>Remove handpieces and purge all waterlines with air to remove water.</li> <li>Remove bottle and clean it thoroughly with hot water, a suitable cleanser and bottle brush.</li> </ol> <p><i>Routine Treatment:</i></p> <ol style="list-style-type: none"> <li>Every week or as needed*, add one (1) 1.7 gram VistaTab Dental Waterline Treatment Tablet for every 750 ml of potable water. Allow the tablet to dissolve.</li> <li>Flow about half of the treated water through the dental waterlines and let the treated water sit within the lines for 5 minutes. Purge the remaining water through the lines.</li> <li>Flush at least 500 ml of potable water through the lines to remove any residual chemical.</li> </ol> <p><i>Note: Periodic cleaning may be done on any day of the week. *Treatment should be done if bacteria count is greater than 500cfu/ml as recommended by the CDC.</i></p> <ol style="list-style-type: none"> <li>Go to “Daily” protocol to prepare a fresh supply of water for use with patients.</li> </ol>
Bottle Capacity	Min Drops	Max Drops																	
500 ml	1	2																	
750 ml	1	3																	
1.0 Liter	1	3																	
1.5 Liters	1	4																	
2.0 Liters	2	5																	

\*\* Note: A line cleaning may be necessary more often than once per month depending on the quality of the water used to fill the bottles, contamination by users during bottle changes, age of system, etc. Be certain to purge all devices that contain water including handpiece, air/water syringe, scaler, quick disconnect line and cup filler (if present). Water quality is greatly impacted by cross connections, dead legs, faulty anti-retraction valves, etc. We suggest testing water quality at least quarterly using an in-office test or outside service. When used properly, VistaClean is completely non-toxic, safe for the environment, non-corrosive, simple to use and has no negative effect on bonding. Never use full strength.

**Precautions:** Use these products only as directed. Please read the *Precautionary Statements* on the reverse side of this page. Go to [www.hu-friedy.com](http://www.hu-friedy.com) for more information on these and other Hu-Friedy products.



Hu-Friedy Mfg. Co., Inc.  
3232 N. Rockwell St. • Chicago, IL 60618  
1-800-HU-FRIEDY • [www.hu-friedy.com](http://www.hu-friedy.com)

Re-order Numbers	
<i>Description</i>	<i>Part Code</i>
Team Vista	IMS-1450
VistaTab	IMS-1451
VistaClean	IMS-1452

**PRECAUTIONARY STATEMENTS**

Hazards to Humans and Domestic Animals

**DANGER.** Corrosive. Causes irreversible eye damage. Harmful if swallowed. Do not get in eyes or on clothing. Wash thoroughly with soap and water after handling and before eating, drinking, chewing gum, or using tobacco. Remove contaminated clothing before reuse.

**FIRST AID**

**If in Eyes:** Hold eye open and rinse slowly and gently with water for 15-20 minutes. Remove contact lenses, if present, then continue rinsing eye. Call a poison control center or doctor for treatment advice.

**If Swallowed:** Call poison control center or doctor for treatment advice. Have person sip a glass of water if able to swallow. Do not induce vomiting unless told to do so by the poison control center or doctor. Do not give anything by mouth to an unconscious person.

Have the product container or label with you when calling poison control center or doctor, or going for treatment.

**NOTE TO PHYSICIAN**

Probable mucosal damage may contraindicate the use of gastric lavage.

**ENVIRONMENTAL HAZARDS**

This product is toxic to fish and other aquatic organisms.

**PHYSICAL/CHEMICAL HAZARDS**

Keep away from combustible, organic or other readily oxidizable materials. Do not mix with any other chemicals.

VistaTab™

**Dental Waterline Cleaner Tablets**

**IMS-1451**

**For the Cleaning and Control of Microbial Contaminants in Dental Unit Waterlines**

*Active Ingredients:*

- Sodium Chlorite ..... 4.0%
- Sodium Dichloroisocyanurate dihydrate . 1.0%
- Other Ingredients:*
- Total: ..... 95.0%
- ..... 100.0%

EPA Reg. No. 70060-25-75372  
EPA Est. No. 73593-OH-001

*\* Sodium chlorite in this product is from Spain*

**KEEP OUT OF REACH OF CHILDREN**

**DANGER**

This box contains: 15 – 1.7 gram tablets  
Net Weight 25.5 grams

Sold by



Hu-Friedy Mfg. Co., Inc.  
3232 N. Rockwell St. • Chicago, IL 60618  
1-800-HU-FRIEDY • www.hu-friedy.com

**DIRECTIONS FOR USE**

It is a violation of Federal law to use this product in a manner inconsistent with its labeling.

**INITIAL TREATMENT:**

1. Add two (2) 1.7 gram *VistaTab* Dental Waterline Treatment Tablets for every 750 ml of potable water. Allow the tablets to dissolve.
2. Flow about half of the treated water through the dental waterlines and let the treated water sit within the lines for 5 minutes. Purge the remaining water through the lines.
3. Repeat steps 1 and 2.
4. Flush at least 500 ml of potable water through the lines to remove any residual chemical.

**ROUTINE TREATMENT:**

1. Every week or as needed\*, add one (1) 1.7 gram *VistaTab* Dental Waterline Treatment Tablet for every 750 ml of potable water. Allow the tablet to dissolve.
2. Flow about half of the treated water through the dental waterlines and let the treated water sit within the lines for 5 minutes. Purge the remaining water through the lines.
3. Flush at least 500 ml of potable water through the lines to remove any residual chemical.

Note: Periodic cleaning may be done on any day of the week. \* Treatment should be done if bacteria count is greater than 500 cfu/ml as recommended by the CDC.

**STORAGE AND DISPOSAL**

Do not contaminate water, food, or feed by storage and disposal. Store in a cool, dry area away from sunlight and heat and in an area inaccessible to children. Keep product from contacting air prior to use. Place empty foil pouch in trash.